



## Sensors for dyb electronic torque meter

**DYC-100N**

DYC-...N

### SAM Outillage

10 rue Camille de Rochetaillée  
42000 Saint-Etienne - France

RCS Saint-Etienne B 338 002 231 | SAS au capital de 8 189 583 €

<https://www.sam-outillage.com/>

Photos and texts not contractually binding. Do not litter. Document generated on 2024-04-26 04:36:52





## DESCRIPTION

### Sensors for dyb electronic torque meter

TO BE USED WITH TORQUEMETER DYB Very high precision torque sensor designed for metrology laboratories. Temperature compensation integrated directly in the measurement gauges. The sensors to be mounted on BC-5N type inspection benches. Characteristics: - Impedance: 2,000 ohms. - Sensitivity: 1.5 mV/ V (full scale). - Power supply: 4 VDC.

## STANDARDS / DIRECTIVES

NF ISO 1174-1, ISO 6789

### SAM Outillage

10 rue Camille de Rochetaillée  
42000 Saint-Etienne - France

RCS Saint-Etienne B 338 002 231 | SAS au capital de 8 189 583 €

<https://www.sam-outillage.com/>

Photos and texts not contractually binding. Do not litter. Document generated on 2024-04-26 04:36:52



**SPECIFIC DATA**

|                 |  |
|-----------------|--|
| Designation     | SENSOR FOR TORQUEMETER DYB 100 NM 3/8" |
| Sales reference | DYC-100N                               |
| Weight (kg)     | 0.38                                   |
| Capacity (Nm)   | 10 - 100                               |
| Dimension (mm)  | 57x44x43                               |
| female square   | 3/8                                    |
| Guarantee       | M                                      |

**Guarantee applied**

**M1 | SAM Equipment and Machine Guarantee:**

Guarantee for a limited period, covering technical tools, devices and machines used under normal conditions. The length is given by the number after the letter M: 1 to 5 years

[>> Find out more](#)

**SAM Outillage**

10 rue Camille de Rochetaillée  
42000 Saint-Etienne - France

RCS Saint-Etienne B 338 002 231 | SAS au capital de 8 189 583 €

<https://www.sam-outillage.com/>

